

24W POWER SUPPLY

The E2 family of AC/DC switch mode power supplies offers an unmatched degree of flexibility. A variety of housings and customization options allow the same reliable electronic core to be used in a wide range of applications.

Features

- Ultra-low standby losses
- High Efficiency
- Protection class II
- Wide selection of output plugs
- Premium quality Japanese brand capacitors
- Manufacturing according to ISO 9001
- Short circuit proof
- Designed in Austria
- Made in the Czech Republic

| Specification | | | | | | |
|-------------------------|-----------|----|--|--|--|--|
| Output Power | 24 | W | | | | |
| Output Voltage | 5 - 24 | V | | | | |
| Output current | 3 | Α | | | | |
| Universal input voltage | 90 - 264 | V | | | | |
| Operating temperature | 0 - 40 | °C | | | | |
| Efficiency | typ. 86,6 | % | | | | |
| Standby Power | typ. 35 | mW | | | | |
| Efficiency level | VI | | | | | |
| Means of protection | 2 x MOPP | | | | | |
| Insulation of output | SELV | | | | | |
| Leakage current | max. 100 | μΑ | | | | |

Approvals













Options

- Power on LED
- Customer specific connectors and housing

| Housing versions | | | | | | |
|--|----|----|-----|--|--|--|
| Wall plug-in, fixed or interchangeable | | | | | | |
| • • | | | 0 0 | | | |
| EU | UK | US | AUS | | | |

| Te | est standards |
|--|-------------------------------------|
| EN 55014-1 EN 55014-2 EN 55032 EN 55024 EN 60601-1-2 | General EMC standards |
| EN 60950-1 UL 60950-1 EN 62368-1 UL 62368-1 | Information technology equipment |
| EN 60335-1 EN 61558-2-16 EN 61558-1 | Household devices |
| EN 60601-1 EN 60601-1-11 ES 60601-1 | Medical electrical equipment |
| EN60601-1-11 | Home healthcare |
| Degree of | environment |
| protection: IP22 | Only for housing types: E, U, G |





| Parameter | Symbol | Min | Тур. | Max | Unit | Test Cond. |
|---|---|----------------|------------|----------------|----------------------------------|--|
| | ations are sul | | | | | |
| | | , | | | | |
| | U _{IN} | 90 | | 264 | V_{AC} | |
| Input Voltage | Opei | | | | voltage may ca s not meet the | |
| Input Current | I _{IN} | 9 | 300 | 800 | mA | |
| Input Frequency | f _{IN} | 47 | 50 | 63 | Hz | |
| Efficiency | η | | 86,6 | | % | at full load |
| Stand-by power | P _{stb} | | 35 | 75 | mW | without load |
| International efficiency mark | | | VI | | | |
| Output Power | P _{out} | | | 24 | W | |
| Output Voltage | U_out | 5 | | 24 | V_{DC} | |
| Output voltage tolerance | Δ Uout PCB | | | 3 | % | at PCB |
| Output voltage tolerance at end of standard cable | Δ Uout cable | | | +3/-5 | % | 12-24V U _{out} 1,5m/0,5mm ² |
| Ripple Voltage | U _{r rms} | | | 50 | mV_{rms} | 1,0111, 0,011111 |
| Output Current | lout | | | 3 | A | |
| Max. Overload current | lout overload | | 140 130 | | % of I _{out} | U _{IN} = 264V U _{IN} = 90V |
| | Maxim | num 1 minute o | | on, followed b | y 15 minute co | |
| Isolation | | | | | oltage (SELV) | |
| Means of protection | | | 2 x N | 10PP | | |
| | Standard | 3 | | | | 50Hz |
| Dielectric Strength | Household | 3,9 | | | kV_AC | sinusoidal |
| | Medical | 4,4 | | | | waveform |
| Leakage current | I _{LK} | | | 100 | μΑ | |
| Operating Temperature | | | | | | |
| 0 → 40°C | T _{OP} | 0 | | 40 | °C | free convection |
| Storage Temperature | | | | | | |
| -30 → 80 °C | T _{ST} | -30 | 25 | 80 | °C | |
| Humidity | | | | | | |
| 95% | | | | 95 | % | non condensing |
| Atmospheric Pressure | | | | | | |
| 70 - 106 kPa | | 70 | | 106 | kPa | |
| Single component failure | A single component failure does not cause any damage to persons or ambient (fire, explosions, etc). | | | | | |
| Disconnecting device | Direct plug-in The power supply itself is the disconnecting device. Socket-outlet shall be easily accessible. | | | | ing device. | |




| | Ordering information and part number example | | | | | | | | |
|----|---|---|-----------------------------|---|---------------|----------------|----|-------------------|------------------------------|
| E2 | С | F | М | W | 3 | L | 24 | 12V | 2,5A |
| | Housing | | Application | | Toleranc e | Options | | Voltage | Current |
| | E Euro plug U US, Canada plug | | S Information Technology | | 3 3% 1 1% | L Power On LED | | 5 - 24V | 0 - 3A |
| | G United Kingdom plug C Interchangeable plug | | H Household M Medical | | on PCB | | | Fixed voltage | |
| | | | | | | | | set at factory | 24W/U _{out} max. |

For versions with output cable please also specify the plug and cable length.

Please refer to the document "secondary plug overview" for our range of standard cables, or specify a customer specific connector as required.

| Reliability | | | | | | |
|---|-------------------|-----------------------|--|--|--|--|
| MTBF | 22,2 years | at 50°C ambient | | | | |
| MTBF calculation according to standards | MIL-HDBK-217 F; - | Notice 1; - Notice 2 | | | | |
| Maintainability | The power supply | is not to be repaired | | | | |







A shorted output does not cause any damage, and normal operation will resume once the short is removed.

| Laser marking | Marking plate symbol explanation | | | | | |
|--|----------------------------------|--|--|--|--|--|
| | C€ | Conformity with the relevant EU directives. | | | | |
| Product name | 22 | ENEC is the high quality European Mark for electrical products that demonstrates compliance with European standards (EN). | | | | |
| Input parameters Output parameters Safety instructions Date code of production CE marking Approval marks | c y us | NRTL Canada / USA Mark issued by Curtis Straus. | | | | |
| | RoHS conform | The power supply has to be disposed appropriately according the local regulations for Waste Electrical and Electronic Equipment. | | | | |
| | | For indoor use only. | | | | |
| | []i | Read instruction manual. | | | | |


| Certification overview | | | | | | |
|------------------------|------------------------|--------------|-----------|--|--|--|
| Housing | Information Technology | Household | Medical | | | |
| EU, UK | CE Z22 IECEE | CE ZZ2 IFGEE | C E FERRE | | | |
| US, Canada | CE cous | CE cous | CE cous | | | |
| Interchangeable Plug | | | CE COUS | | | |
















| Packaging and weight | | | | | | |
|---|------|------|---------------|--|--|--|
| E2EFSW3 24 LV6 | pcs | kg | size (mm) | | | |
| Single Carton | 1 | 0,16 | 125x67x44 | | | |
| Power Supply per Packaging Case | 50 | 8 | 371x266x242 | | | |
| Power Supply per Layer (EU- Pallet) 9 Packaging cases | 450 | 93 | 1200x800x242 | | | |
| 1 Full Pallet (6 Layer) | 2700 | 453 | 1200x800x1500 | | | |
| E2GFSW3 24 LV6 | pcs | kg | size (mm) | | | |
| Single Carton | 1 | 0,18 | 135x74x49 | | | |
| Power Supply per Packaging Case | 50 | 9 | 406x286x272 | | | |
| Power Supply per Layer (EU- Pallet) 8 Packaging cases | 400 | 93 | 1200x800x272 | | | |
| 1 Full Pallet (5 Layer) | 2000 | 360 | 1200x800x1500 | | | |
| E2UFSW3 24 LV6 | pcs | kg | size (mm) | | | |
| Single Carton | 1 | 0,16 | 125x67x44 | | | |
| Power Supply per Packaging Case | 50 | 8 | 371x266x242 | | | |
| Power Supply per Layer (EU- Pallet) 9 Packaging cases | 450 | 93 | 1200x800x242 | | | |
| 1 Full Pallet (6 Layer) | 2700 | 453 | 1200x800x1500 | | | |
| E2CFSW3 30 LV6 | pcs | kg | size (mm) | | | |
| Single Carton (including Power Supply and 4 Adapters) | 1 | 0,24 | 210x74x50 | | | |
| Power Supply per Packaging Case | 25 | 6 | 406x286x272 | | | |
| Power Supply per Layer (EU- Pallet) 8 Packaging cases | 200 | 69 | 1200x800x272 | | | |
| 1 Full Pallet (5 Layer) | 1000 | 261 | 1200x800x1500 | | | |



| EMC - Special requirements according medical standard (Only for medical devices) | | | | | |
|--|--|--|--|--|--|
| Intended use and intended environment | Home healthcare and/or Professional environment | | | | |
| Basic safety and essential performance of the EUT | The power supply unit is not a medical end product, therefore no essential performance is defined by the manufacturer. | | | | |
| Basic safety regarding EMC | The power supply has to ensure proper output voltage according to its characteristics, without service within expected service life. | | | | |
| | Medical electrical equipment needs special precautions regarding EMC and needs to be installed according to EMC information. | | | | |
| | PE of power supply shall be connected to PE of end medical product. User shall not modify power supply. | | | | |
| WARNINGS | The switch mode power supply is designed to achieve the EMI behavior of the specified environment, it includes specific EMI filter to reduce the emissions which are specified in the IEC60601-1-2 standard. | | | | |
| | Please read the complete technical documentation to avoid adverse events to the patient and operator. Read also instructions for use. | | | | |

EMC - Environment

The power supply is intended for use in the electromagnetic environment specified below. The customer or the user of the power supply should assure that it is used in such an environment.

| Emissions test | Compliance | | Electromagnetic enviror | nment - guidance | | | |
|---|---|---|---|--|--|--|--|
| RF emissions CISPR 11 | Group 1 | | The power supply uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. | | | | |
| RF emissions CISPR 11 | Class B | | | | | | |
| Harmonic emissions IEC 61000-3-2 | Complies | | The power supply is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply | | | | |
| Voltage fluctuations / flicker emissions IEC 61000-3-3 | Complies | network that supplies buildings used for domestic purposes. | | | | | |
| Immunity test | EN 60601-1-2:2 test level | 2015 | Achieved levels according EN 60601-1-2:2015 and achieved levels from additional standards. | Electromagnetic environment - guidance | | | |
| Electrostatic | ± 8 kV contact | | ± 8 kV contact | Floors should be wood, concrete or | | | |
| discharge (ESD) IEC 61000-4-2 | ±2 kV, ± 4 kV, ± ± 15 kVair | 8 kV, | ±2 kV, ± 4 kV, ± 8 kV, ± 15 kVair | ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%. | | | |
| Electrical fast transient/burst IEC 610004-4 | ± 2 kV 100 kHz repetition frequency | | ± 2 kV (mains input), 100 kHz ± 2 kV (DC output), 5 kHz | Mains power quality should be that of a typical commercial or hospital environment. | | | |
| Surge IEC 61000-4-5 | Line-Line:± 0,5 k kV Line-to-ground: kV, ± 1 kV, ± 2 | ± 0 ,5 | ±1 kV symmetrical – Differential mode (AC), ±2 kV symmetrical – Common mode (AC), ±0.5 kV symmetrical – | Mains power quality should be that of a typical commercial or hospital environment. | | | |





| | | Differential mode (DC), ±0.5 kV symmetrical – Common mode (DC), 1.2/50 us Open Circuit Voltage | |
|---|--|--|---|
| Voltage dips, short interruptions | 0 % Ut; 0,5 cycle At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° | 0 % Ut; 0,5 cycle At 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° | Mains power quality should be that of a typical commercial or hospital |
| and voltage variations on power supply input lines IEC 61000-4-11 | and 70 % Ut; 25/30 cycles | 0 % Ut; 1 cycle and 70 % Ut; 25/30 cycles Single phase: at 0° | environment. If the user of the power supply requires continued operation during power mains interruptions, it is recommended that the power supply is powered from an uninterruptible power supply or battery. |
| Power frequency (50/60 Hz) magnetic field IEC 61000-4-8 | 30 A/m | 1, 3, 30 A/m | Power should be at levels characteristic of frequency magnetic fields a typical location in a typical commercial or hospital environment. |
| Conducted RF IEC 61000-4-6 | 6 Vrms 150 kHz to 80 MHz | 6 Vrms | Portable and mobile RF communications equipment should not be used closer to any part of the power supply, including cables, than the recommended separation distance. |
| Radiated RF IEC 61000-4-3 | 10 V/m 80 MHz to 2.7 GHz | 10 V/m | Recommended separation distances see following table. |

Field strengths from fixed transmitters such as base stations for radio (cellular/cordless) telephones, land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast, cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters an electromagnetic site survey should be considered. If the measured field strength in the location in which the power supply is used, exceeds the applicable RF compliance level above, the power supply should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the power supply.

Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey should be less than the compliance level in each frequency range. Over the frequency range 150 kHz to 80 MHz, field strength should be less than 3 V/m.

Interference may occur in the vicinity of equipment marked with the following symbol:

| | Frequency range and Level: RF wireless communication equipment | | | | | |
|--|--|----------------------------------|-------------------------|--|--|--|
| Proximity fields from RF wireless communications equipment IEC 61000-4-3 | Test Frequency (MHz) | Modulation | Immunity Level (V/m) | Supplementary information: | | |
| | 385 | **Pulse Modulation: 18Hz | 27 | EUT powered at one of the nominal | | |
| | 450 | *FM ±5Hz deviation: 1kHz sine | 28 | input voltages and frequencies. | | |
| | 710 745 780 | **Pulse Modulation: 217Hz | 9 | Dwell time minimum 1s. Actual dwell time noted in results table. Note * - As an alternative to FM | | |
| | 810 870 930 | **Pulse Modulation: 18Hz | 28 | modulation, 50% pulse modulation at 18Hz may be used because while it does not represent actual modulation, it would be worst case. | | |
| | 1720 1845 1970 | **Pulse Modulation: 217Hz | 28 | Note ** - The carrier shall be modulated using 50% duty cycle | | |
| | 2450 | 2450 **Pulse Modulation: 217Hz | | square wave signal. | | |
| | 5240 5500 5785 | **Pulse Modulation: 217Hz | 9 | | | |



Recommended separation distances between portable and mobile RF communications equipment and the power supply

The power supply is intended for use in the electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the power supply can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the power supply as recommended below, according to the maximum output power of the communication equipment.

| Rated maximum output | Separation distance according to frequency of transmitter (m) | | | | |
|----------------------|---|-------------------|--------------------|--|--|
| power of transmitter | 150 kHz to 80 MHz | 80 MHz to 800 MHz | 800 MHz to 2.5 GHz | | |
| (W) | d = 1.2√P | d = 1.2√P | d = 2.3√P | | |
| 0.01 | 0.12 | 0.12 | 0.23 | | |
| 0.1 | 0.38 | 0.38 | 0.73 | | |
| 1 | 1.2 | 1.2 | 2.3 | | |
| 10 | 3.8 | 3.8 | 7.3 | | |
| 100 | 12 | 12 | 23 | | |

For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be determined using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.

NOTE 4 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.



Energy Efficiency

This power supply family fulfills Directive 2009/125/EC with Commission Regulation (EU) 2019/1782. The vales "Average active efficiency", "Efficiency at low load" and "No-load power consumption" are typical measured values, measured at one representative sample at an input voltage of 230VAC.

| Input specification | | | | |
|---------------------|---------|-----|--|--|
| Input Voltage | 100-240 | VAC | | |
| Input Frequency | 50-60 | Hz | | |

| | | | | Output | specificati | on |
|--|-------|-------|-------|--------|-------------|-----|
| Output voltage | 5 | 9 | 12 | 15 | 24 | VDC |
| Output current | 3 | 2,67 | 2 | 1,6 | 1 | Α |
| Output power | 15 | 24 | 24 | 24 | 24 | W |
| Average active efficiency (100%/75%/50%/25%) | 82,39 | 86,75 | 87,90 | 87,45 | 88,46 | % |
| Efficiency at low load (10 %) | 80,28 | 83,63 | 82,98 | 83,30 | 81,48 | % |
| No-load power consumption | 22 | 32 | 33 | 36 | 49 | mW |

| Revision | Date | Author | Change |
|----------|------------|---------|---|
| А | 01.09.2016 | Krimmel | First edition |
| В | 08.03.2017 | Mauritz | Disconnecting Device added |
| С | 07.08.2017 | Mauritz | HV testing voltage changed |
| D | 08.08.2017 | Mauritz | Reliability, EMC added, Approval and test standards |
| Е | 19.12.2017 | Mauritz | MTBF added |
| F | 30.03.2018 | Krimmel | Update to new document design |
| G | 06.03.2019 | Mauritz | ENEC at medical devices removed |
| Н | 21.03.2019 | Mauritz | Disconnecting device changed |
| I | 27.08.2019 | Mauritz | Test standards changed |
| J | 04.02.2020 | Mauritz | Energy Efficiency added |
| K | 14.04.2020 | Mauritz | Parameter Symbols added, Test standards changed |

CONFIDENTIAL

This document contains proprietary information originated and/or owned by EGSTON System Electronics Eggenburg GmbH. This information shall not be duplicated, used or disclosed in whole, or in part, to any other party or used for any other purpose without the prior consent of EGSTON System Electronics Eggenburg GmbH.

> Copyright © 2020 EGSTON System Electronics Eggenburg GmbH, A-3730 Eggenburg, Grafenberger Straße 37 All Rights Reserved.

www.exista.ch